



Product Catalog
and Surgical Technique



Eztetic[®] 3.1 mmD

Dental Implant System



 **ZimVie**

ZimVie DENTAL SOLUTIONS

Table of Contents

Overview	1
General Information	1
Implant Design and Specifications	3
Crestal Options	3
Conical, Double Friction-Fit Technology	3
Implant Platform	4
Implant Placement Guidelines	4
Indications for Use	4
Eztetic Dental Implant System	5
Implants	5
Prosthetics (NP Platform)	6
• Healing Components	6
• Impression Transfer Components	6
• Provisional Restoration Components	7
• Cement-Retained Components	7
• "Cast To" Gold Abutment Components	7
• OverdenSURE® Removable Attachment System	8
• Prosthetic Accessories and Tools	8
Eztetic Instrumentation	9
Surgical Kit	9
NP Surgical Module	10
Color Coding	11
NP Surgical Module Ordering Information	11
Replacement Instrumentation Ordering Information	11
Preoperative Planning	12
Surgical Technique	14
Drilling Sequence Guidelines	14
Surgical Procedure	15
References	16

Overview


General Information

The 3.1 mmD Eztetic Dental Implant offers a narrow, powerful solution for demanding anterior spaces. This innovative dental implant is designed to provide initial stability¹, allowing immediate restorations when clinically appropriate. The conical, Double Friction-Fit™ Connection with platform switch, combined with narrow platform (NP) prosthetics, are designed for crestal bone maintenance² and good aesthetics by accommodating maximum soft-tissue volume. The NP Surgical Module conveniently snaps into the Tapered Screw-Vent® Surgical Kit for streamlined surgical procedures.



Restorative Profile for Aesthetic Emergence:

Implant-abutment connection along with a contour abutment profile are designed to provide space for soft-tissue and aesthetic emergence of the restoration





Strength³ for Long-Lasting Aesthetics

Implant design and a conical Double Friction-Fit Connection are combined for exceptional strength, reduced micromovement, and microleakage

Primary Stability¹ for Immediate Aesthetics

Tapered implant geometry combined with dedicated soft- and dense-bone surgical protocols are designed for high primary stability in all types of bone



MTX[®] Microtexture
at 2000x Magnification

MTX Surface for Increased Bone Apposition^{3,4}

The MTX Microtextured Surface has been documented to achieve high levels of bone-to-implant contact and successful clinical results under conditions of immediate loading


Coronal Options for Bone Level Maintenance

The coronal microgrooves are designed to preserve crestal bone.⁵
Two coronal surface configurations are available:

- Full MTX Microtexturing with MTX Crestal Microgrooves (Model CT)
- 0.5 mm Machined Collar with MTX Crestal Microgrooves (Model CM)

Implant Design and Specifications

Crestal Options: Fully Textured with Microgrooves and 0.5 mm Machined Collar with Microgrooves




The 3.1 mmD Eztetic Dental Implants have an MTX microtextured or 0.5 mm machined coronal aspect, followed by 1.8 mm of the MTX Surface with microgrooves. The six microgrooves are circumferential with a depth of 0.06 mm and peak-to-peak width of 0.3 mm. Triple-lead threads begin immediately after the microgrooves and continue to the apex. The degree of body taper varies between 1° and 4°, depending on implant length.

Conical, Double Friction-Fit Technology

The 3.1 mmD Eztetic Dental Implants feature a conical Double Friction-Fit Connection designed to reduce micromovement and microleakage via a precise implant-abutment interface, aiding in crestal bone maintenance.


Figure 1A The implant-abutment connection offset along with a Contour Abutment are designed to provide space for soft-tissue and aesthetic emergence of the restoration.

Figure 1B The 3.1 mmD Eztetic Dental Implants feature a conical Double Friction-Fit Connection designed to reduce micromovement and microleakage via a precise implant-abutment interface, aiding in crestal bone maintenance.




Implant Platform

The implant platform diameter is measured across the most coronal part of the implant. The 3.1 mmD Eztetic Dental Implant features a 2.9 mmD prosthetic platform. A 1.3 mm deep, 17° internal cone extends from the outermost diameter (2.9 mmD) of the implant platform to the internal hex of the implant. The internal hex is 2.1 mm flat-to-flat with a depth of 1.7 mm (Fig. 2A and 2B). The 3.0 mm deep conical connection is designed to distribute stresses deep into the implant and away from the crestal bone to aid in crestal bone maintenance.



Implant Placement Guidelines

For Pre-Surgical Planning Instructions and Anatomical Criteria guidelines please refer to the Tapered Screw-Vent Surgical Manual.



Collar Depth Adjustment

The 3.1 mmD Eztetic Dental Implants are designed to be placed at bone level or slightly below bone level, according to clinical preference and patient anatomy.

Indications for Use

The 3.1 mmD Eztetic Dental Implants are designed for use in the anterior maxilla or mandible for immediate loading or for loading after a conventional healing period. Implants may be used to replace one or more missing teeth. Immediate loading is indicated when there is good primary stability and an appropriate occlusal load.

The 3.1 mmD Eztetic Dental Implants may be placed immediately following an extraction or loss of natural teeth provided there is sufficient volume of alveolar bone to minimally support the implant (minimum 1.0 mm circumferential and 2.0 mm apical).

The 3.1 mmD Eztetic Dental Implants should be splinted to additional implants when used in the pre-molar region and should not be used in the molar region. For more information, please review the full Instructions for Use.




Eztetic Dental Implant System

Implants


Implant Platform Color-Coding

3.1 mmD Eztetic Dental Implants are available with a 2.9 mmD prosthetic platform.

Implant Diameter	Implant Platform	Color-Coding
3.1 mmD	NP (2.9 mmD)	 Light Blue

NP = Narrow Platform

Abutment Emergence Profile Compatibility

Abutment Emergence Profile*	Color-Coding
4.5 mmD	 Tan

*For Contour components.


NOTE: Contour Abutments with 3.7 mmD Emergence Profile are not compatible with 3.5 mmD Contour Restorative Components.

in NL geen
voorraadartikel



Eztetic Dental Implants, MTX Surface, Fully Textured with Microgrooves


Includes Healing Screw (CCSNP)

Implant Diameter	Implant Platform	Implant Length				
		8.0 mmL	10 mmL	11.5 mmL	13 mmL	16 mmL
3.1 mmD	 NP (2.9 mmD)	CT318	CT3110	CT3111	CT3113	CT3116




Eztetic Dental Implants, MTX Surface, 0.5 mm Machined Collar with Microgrooves

Includes Healing Screw (CCSNP)

Implant Diameter	Implant Platform	Implant Length				
		8.0 mmL	10 mmL	11.5 mmL	13 mmL	16 mmL
3.1 mmD	 NP (2.9 mmD)	CM318	CM3110	CM3111	CM3113	CM3116

Implant Fatigue Strength¹


The 3.1 mmD Eztetic Dental Implants achieved 43% higher fatigue strength compared to selected competitive implants of similar diameters.¹



All Products were tested in increments of 5.

Insertion Torque¹


The 3.1 mmD Eztetic Dental Implants achieved high insertion torque.¹



Benchtop engineering test utilizing a dense bone substrate.¹

Torsional Yield Strength¹

The 3.1 mmD Eztetic Dental Implant interface withstood higher torsional forces than the selected competitors.¹



Benchtop engineering test utilizing the implants and their corresponding drivers.¹

Eztetic Dental Implant System

Prosthetics (NP Platform)

Healing Components



Surgical Healing Screw, Replacement

Implant Platform	Item No.
NP (2.9 mmD)	CCSNP

Healing Collars



Implant Platform	Emergence Profile	Cuff Height		
		1.5 mm	3.0 mm	4.5 mm
NP (2.9 mmD)	3.7 mmD	CHCNP31	CHCNP33	CHCNP34
NP (2.9 mmD)	4.5 mmD	CHCNP41	CHCNP43	CHCNP44

Figure A



NP Platform

Figure B

The top surface of the healing collar is etched with three symbols to reference implant platform diameter (left), emergence profile diameter (top right) and cuff height (lower right). In some cases, only the initial digit of the measurement is shown. See item number table above for specific measurements. NP = Narrow Platform (2.9 mmD)



Impression Transfer Components

Indirect Transfers (Closed-Tray Procedure)

Color-coded by implant platform. Includes a Retaining Screw (CASLT)



Implant Platform	Emergence Profile	Cuff Height		
		1.5 mm	3.0 mm	4.5 mm
NP (2.9 mmD)	3.7 mmD	CITNP31	CITNP33	CITNP34
NP (2.9 mmD)	4.5 mmD	CITNP41	CITNP43	CITNP44
Replacement Retaining Screw		CASLT	CASLT	CASLT

Direct Transfers (Open-Tray Procedure)

Color-coded by implant platform. Includes a Retaining Screw (CASLC)



Implant Platform	Emergence Profile	Cuff Height		
		1.5 mm	3.0 mm	4.5 mm
NP (2.9 mmD)	3.7 mmD	CDTNP31	CDTNP33	CDTNP34
NP (2.9 mmD)	4.5 mmD	CDTNP41	CDTNP43	CDTNP44
Replacement Retaining Screw		CASLC	CASLC	CASLC



Implant Analog, Titanium

Color-coded by implant platform.

Implant Platform	Item No.
NP (2.9 mmD)	CIANP

Provisional Restoration Components

Titanium Temporary Abutment

Includes a Retaining Screw (CUAS) and a Long Processing Screw (CASLC)



Implant Platform	Emergence Profile	Item No.
NP (2.9 mmD)	3.5 mmD	CTANP31
Replacement Retaining Screw		CUAS
Replacement Long Processing Screw		CASLC

Cement-Retained Components

Contour Abutments, Straight

Includes a Retaining Screw (CUAS)



Implant Platform	Emergence Profile	Cuff Height		
		1.5 mm	3.0 mm	4.5 mm
NP (2.9 mmD)	3.7 mmD*	CANP31S	CANP33S	CANP34S
NP (2.9 mmD)	4.5 mmD**	CANP41S	CANP43S	CANP44S
Replacement Retaining Screw		CUAS	CUAS	CUAS

Contour Abutments, Angled 17°

Includes a Retaining Screw (CUASA or CUAS)



Implant Platform	Emergence Profile	Cuff Height	
		1.5 mm	3.0 mm
NP (2.9 mmD)	3.7 mmD*	CANP31A	CANP33A
NP (2.9 mmD)	4.5 mmD**	CANP41A	CANP43A
Replacement Retaining Screw		CUASA	CUAS

Angled Abutments, 20°

Includes a Retaining Screw (CUAS)



Implant Platform	Emergence Profile	Item No.
NP (2.9 mmD)	3.5 mmD	C20A3
Replacement Retaining Screw		CUAS

"Cast To" Gold Components

"Cast To" Gold Abutments, Engaging

Includes a Retaining Screw (CUAS)



Implant Platform	Emergence Profile	Item No.
NP (2.9 mmD)	3.5 mmD	CEANP31
Replacement Retaining Screw		CUAS
Long Processing Screw		CASLC

* Contour Abutments with 3.7 mmD Emergence Profile are not compatible with 3.5 mmD Contour Restorative Components.

** Please refer to the Tapered Screw-Vent Implant System Catalog for a list of 4.5 mmD Contour Restorative Components compatible with the 4.5 mmD Emergence Profile.

OverdenSURE® Removable Attachment System

OverdenSURE Abutments

U.S. only



Implant Platform	Collar Height					
	1.0 mm	2.0 mm	3.0 mm	4.0 mm	5.0 mm	6.0 mm
NP (2.9 mmD)	ODSA-NP201	ODSA-NP202	ODSA-NP203	ODSA-NP204	ODSA-NP205	ODSA-NP206

Retention Kit
ODSZ-RIKIT2001 not included,
please order separately.

All markets outside the U.S.



Implant Platform	Collar Height					
	1.0 mm	2.0 mm	3.0 mm	4.0 mm	5.0 mm	6.0 mm
NP (2.9 mmD)	ODS-NP201	ODS-NP202	ODS-NP203	ODS-NP204	ODS-NP205	ODS-NP206

Retention Kit ODSZ-RIKIT2001
included in all active
geographies except the U.S.

OverdenSURE Retention Inserts

U.S. only



Description	Qty	Item No.
Retention Kit, 20° Includes red, blue, pink, clear retention inserts and a retention housing cap	1 set	ODSZ-RIKIT2001
	2 sets	ODSZ-RIKIT2002
Retention Kit 40° Includes a red, yellow, green retention inserts and a retention housing cap.	1 set	ODSZ-RIKIT4001
	2 sets	ODSZ-RIKIT4002

All markets outside the U.S.

Description	Qty	Item No.
Retention Kit, 20° Includes red, blue, pink, clear retention inserts and a retention housing cap	1 set	ODSZ-RIKIT2001
	2 sets	ODSZ-RIKIT2002
Retention Kit 40° Includes a red, yellow, green retention inserts and a retention housing cap.	1 set	ODSZ-RIKIT4001
	2 sets	ODSZ-RIKIT4002

Please refer to the OverdenSURE Brochure [ZV0812US] for additional options.

OverdenSURE Components and Tools



Discription	Qty	Item No.
Analog	1	ODS-AA01
	4	ODS-AA04
	10	ODS-AA010



Discription	Qty	Item No.
Abutment	1	ODS-BARATT01
Bar	2	ODS-BARATT02
Attachment	4	ODS-BARATT04



Discription	Qty	Item No.
Retention Housing Cap	1	ODS-RH01
	4	ODS-RH04
	10	ODS-RH010



Discription	Qty	Item No.
Impression Coping	1	ODS-IC01
	4	ODS-IC04
	10	ODS-IC010



Discription	Qty	Item No.
Blackout Space	20	ODS-BS020
Abutment Driver	1	ODS-DRVR
Removal & Insertion Tool	1	ODS-IRTOOL

Ball Abutment

Abutments do not engage internal hex connection.



Implant Platform	Cuff Height		
	2.0 mm	4.0 mm	6.0 mm
NP (2.9 mmD)	CBANP2	CBANP4	CBANP6

Cap Attachment Housing and
Nylon Liner are sold separately
(CA).

Prosthetic Tools and Tools




Description	Item No.
Hex Driver, Short, with GemLock® Retention (1.25 mm, 22 mmL)	HXGR1.25
Hex Driver, Long, with GemLock Retention (1.25 mm, 30 mmL)	HXLGR1.25
Latch-Lock Hex Driver (Short) for Surgical and Prosthetic Screws (may require Reduction Handpiece) (1.25 mm, 23 mmL)	HX1.25D
Latch-Lock Hex-Driver (Long) for Surgical and Prosthetic Screws (may require Reduction Handpiece) (1.25 mm, 26 mmL)	HXL1.25D
Torque Wrench, Restorative (adjustable torque range 10 Ncm – 35 Ncm)	TWR
Torque Wrench Hex Driver, Short (1.25 mm, 17 mmL)	TW1.25
Torque Wrench Hex Driver, Long (1.25 mm, 22 mmL)	TW1.25L
Removal Tool for Eztetic Abutments	CLRT2

Note: CLRT2 is included with the NP Surgical Module.

Eztetic Instrumentation


Tapered Screw-Vent Surgical Kit



NP Surgical Module



The NP Surgical Module snaps into the Tapered Screw-Vent Surgical Kit



For maximum cutting efficiency, replace drills frequently.

Color Coding

Implant diameter 3.1 mmD

Surgical sequence color bar



Drill band color for dense bone protocol



Implant cap color and restorative platform



Vial cap label

NP: Narrow Platform (2.9 mmD)



NP Surgical Module



Description	Qty.	Item No.
NP Surgical Module (to be inserted into TSVKITG)	1 Ea.	NPMODG
NP Surgical Module (Tray Only)		NPTRAYG
2.4 mmD Dríva™ Drill, 22 mmL		EZT24G
2.8/2.4 mmD Dríva Step Drill, 22 mmL		EZT28D24G
3.0 mmD Cortical Bone Tap Tool		ZOPTT30
2.1 mm GemLock Hex Drill, Latch Lock, Short		CHD2.1
2.1 mm GemLock Hex Drill, Latch Lock, Long		CHDL2.1
2.1 mm GemLock Hex Tool, Ratchet, Short		CHR2.1
2.1 mm GemLock Hex Tool, Ratchet, Long		CHRL2.1
Removal Tool for Eztetic Abutments		CLRT2

Please refer to the Tapered Screw-Vent Implant System Catalog for a complete list of surgical instrumentation in the Instrument Kit System for TSV connections.

Replacement Surgical Instrumentation



Description	Item No.
2.4 mmD Dríva Drill, 22 mmL	EZT24G
2.8/2.4 mmD Dríva Step Drill, 22 mmL	EZT28D24G
3.0 mmD Cortical Bone Tap Tool	ZOPTT30
2.1 mm GemLock Hex Drill, Latch Lock, Short	CHD2.1
2.1 mm GemLock Hex Drill, Latch Lock, Long	CHDL2.1
2.1 mm GemLock Hex Tool, Ratchet	CHR2.1
2.1 mm GemLock Hex Tool, Ratchet	CHRL2.1
Removal Tool for Eztetic Abutments	CLRT2

Preoperative Planning

Preoperative Planning:

Proper treatment planning, as well as the selection of the proper implant length and diameter, are crucial to the long-term success of the implant and restoration.

Before an implant can be selected, the anatomical foundation available to receive the implant must be carefully assessed. Several steps should be taken to complete the evaluation:

1. Clinical examination of the oral cavity can provide important information about the health of the soft tissue at the proposed implant site. Tissue tone and the state of the superficial tissues should be evaluated. In addition, the patient should demonstrate an adequate dimension of attached gingiva or keratinized tissue at the site selected for implantation. In partially edentulous cases, the periodontal status of the remaining dentition should be assessed and interaction between the implant restoration and the adjacent natural dentition should be considered.
2. The bony foundation and ridge need to be clinically analyzed to ensure the presence of proper dimensions and the amount of bone for implant placement. At least one millimeter of bone should be present at the buccal and lingual aspects of the implant following placement. During the planning stage, it is useful to measure the existing bone foundation.

Note: Please ensure as many implants as necessary are used for a fully stable restoration.

CT Scans:

Computed tomography (CT) scans help surgeons view parts of the body with three-dimensional images. Image-guided surgical planning allows surgeons to see anatomical landmarks such as nerves, sinus cavities, and bony structures in order to plan for the placement of dental implants and prostheses.

Through the use of CT scans, clinicians are able to more precisely measure the locations of anatomical structures, dimensions of the underlying bone, and certain bone densities in order to plan and treat clinically demanding cases.

Radiographic Transparencies:

The vertical height of the bone can be determined radiographically. Accurate measurement of the vertical dimension on the radiograph facilitates the selection of the appropriate implant length. This helps to avoid implant placement into the maxillary sinus, the floor of the nose or the mandibular canal, and prevents perforation of the inferior aspect of the mandible. Measurements can be made directly on the panoramic radiograph using a millimeter ruler. Corrections should be made for the degree of enlargement or reduction produced by the particular radiographic equipment.

Radiographic marking balls of a known dimension can be embedded in a plastic template prior to radiographic examination. Once the radiograph is taken and the metal marking balls are visible on the image, measurements can be taken to determine the amount of bone available for implant placement.

To calculate the distortion factor, a simple formula can be utilized: $(5 \div A) \times B$ = the amount of actual bone available.

Formula Key =

- Radiographic marking ball = 5.0 mm in diameter.
- A = Size of marking ball image on radiograph.
- B = Length in millimeters on the radiograph of available bone between the crest of the ridge and the inferior alveolar canal.


Example:

A = 6.5 mm

B = 14 mm

Therefore: $(5 \div 6.5) \times 14 = 10.76$ mm actual bone available

Note: A 2.0 mm margin of safety, from the apical end of the implant to any adjacent vital structure, should be considered.



Preoperative Planning

Radiographic Transparencies Instructional Steps:

A dental implant radiographic transparency supports the preoperative implant treatment planning process. A radiographic transparency is overlaid onto a radiograph to assist the clinician in the preoperative determination of options for implant length and diameter. It is used in conjunction with a 5 mm radiographic marking ball. Representations of the implant and the 5 mm radiographic marking ball are shown on the radiographic transparency at 100%, 115%, and 125% scales.

Visually inspect the transparency before each use for damage. The transparency should not be used if damaged or deteriorated. The following steps outline the proper use of the radiographic transparency in conjunction with the 5 mm radiographic marking ball(s) during preoperative planning:

1. Overlay the 100%, 115%, and 125% scaled 5 mm circular radiograph ball outline found on the transparency over the 5 mm radiographic ball image on the radiograph and determine which outline is closest to the diameter of the radiographic ball image on the radiograph. If the radiographic ball image on the radiograph extends outside the circular border of the radiographic ball outline on the 100% scale, use either the 115% or 125% scale for measurement estimations. If the radiographic ball image extends outside the circular border of the radiographic ball outline on the 125% scale, DO NOT use this radiographic transparency and refer to the Radiographic Marking Balls procedure to determine approximate bone height (see section on calculation of distortion factor on page 12).

Note: The radiographic ball should maintain its spherical shape on the radiograph, otherwise distortion that cannot be measured may have occurred. If this happens, it is recommended that a new radiograph be taken.

2. Select the scale (100%, 115%, or 125%) to use based on which circular radiograph ball outline best matches the diameter of the radiographic ball image on the radiograph.
3. To determine an approximation of available vertical bone height at the proposed implant site, align the zero mark on the selected ruler (100%, 115%, or 125%) to the crest of the edentulous ridge and measure the length between the crest and anatomical structures in the proposed implant site including the floor of the maxillary sinus, the floor of the nose, and the mandibular canal.

Note: A minimum of 2 mm margin of safety, from the apical end of the implant to the adjacent vital structure, should be considered.


4. Overlay the implant silhouette corresponding to the selected scale (100%, 115%, or 125%) onto the proposed implant site to visually estimate if adequate vertical bone height is present for the selected implant length.

Note: The intended use of this device is exclusively for preoperative planning and to be used as a guide. Implant length and diameter should not be determined solely by relying on the radiographic transparency.

Soft bone protocol: follow solid color bars on the surgical tray surface until the segmented color bar. The segmented color bar indicates the final drill for soft bone protocol.

Diagram of the NP Surgical Module showing various ports and connections:

- 2.4mm
- 2.8/2.4mm
- 3.0mm Tap
- 2.1mm Hex
- CLRT2
- NP Surgical Module
- Extra



Surgical Technique

Surgical Procedure

Please follow General Surgical Instructions and Cleaning and Sterilization Guidelines described in the Instructions for Use (IFU) and Tapered Screw-Vent Implant System Surgical Manual [ZBISNT0007].

Site Preparation

Please follow Site Preparation Instructions described in the Tapered Screw-Vent Implant System Surgical Manual [ZBINST0007] along with a drilling sequence for the 3.1 mmD Eztetic Implant.


Removing the Implant from the Vial

Remove the implant outer vial from the box and open the outer vial to break the seal. Drop the sterile inner vial and contents onto a sterile field. Flip the white top of the inner vial open by pressing on the flat side with access hole. Press the top to the inner vial body to lock in the top. Grasp the top of the titanium packaging component placed on top of the implant, remove it, and discard.




Delivering the Implant to the Site

Place the appropriate insertion instrument directly into the implant. The following instruments can be used for implant delivery to the site: the GemLock Hex Drill [CHD2.1, CHDL2.1] attached to a motor handpiece, or GemLock Hex Driver [CHR2.1, CHRL2.1] attached to the GemLock Retaining Square Ratchet [RSR] or Stainless Steel Screwdriver Handle [SSHHS]. Carry the implant via the selected delivery instrument(s) to the receptor site and place directly into the prepared osteotomy.



Inserting and Orienting the Implant

Rotate the implant into place with the selected delivery instrument(s). The GemLock Hex Drills and Drivers are designed with six flats to align with the implant hex. To ensure proper orientation of the Contour Abutments, align the flat side of the Hex Drill or Driver to the buccal aspect. For Angled Abutments, orient a flat side of the Hex Drill or Driver toward the direction of the implant angle. Follow One-Stage or Two-Stage Healing Instructions provided in the Instructions for Use.





References

1. Data on file.
2. Chu C-M, Huang H-L, Hsu J-T, Fuh L-J. Influences of internal tapered abutment designs on bone stresses around a dental implant: three-dimensional finite element method with statistical evaluation. *J Periodontol* 2012;83:111-118.
3. Trisi P, Marcato C, Todisco M. Bone-to-implant apposition with machined and MTX microtextured implant surfaces in human sinus grafts. *Int J Periodontics Restorative Dent*. 2003;23(5):427-437.
4. Todisco M, Trisi P. Histomorphometric evaluation of six dental implant surfaces after early loading in augmented human sinuses. *J Oral Implantol*. 2006;32(4):153-166.
5. Shin SY, Han DH. Influence of a microgrooved collar design on soft and hard tissue healing of immediate implantation in fresh extraction sites in dogs. *Clin Oral Implants Res*. 2010;21:804-814.

For more information, visit [ZimVie.com](https://www.ZimVie.com)

ZimVie Dental
4555 Riverside Drive
Palm Beach Gardens, FL 33410
1-800-342-5454
Phone: +1-561-776-6700
Fax: +1-561-776-1272



www.implacom.nl | +31 (0)577 46 1927



Unless otherwise indicated, as referenced herein, all trademarks and intellectual property rights are the property of ZimVie Inc. or an affiliate; and all products are manufactured by one or more of the dental subsidiaries of ZimVie Inc. (Biomet 3i, LLC, Zimmer Dental, Inc., etc.) and marketed and distributed by ZimVie Dental and its authorized marketing partners. OverdenSURE is manufactured by Terrats Medical S.L. TWR (Torque Wrench) is manufactured by Joset Ganter. For additional product information, please refer to the individual product labelling or instructions for use. Product clearance and availability may be limited to certain countries/regions. This material is intended for clinicians only and does not comprise medical advice or recommendations. Distribution to any other recipient is prohibited. This material may not be copied or reprinted without the express written consent of ZimVie. ZVINST0111



REV A 01/24 ©2024 ZimVie. All rights reserved.